



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/022,466	12/13/2001	Thomas H. Gilman	30056/37712	1716
4743	7590	02/10/2004		
MARSHALL, GERSTEIN & BORUN LLP 6300 SEARS TOWER 233 S. WACKER DRIVE CHICAGO, IL 60606			EXAMINER STASHICK, ANTHONY D	
			ART UNIT 3728	PAPER NUMBER

DATE MAILED: 02/10/2004

5

Please find below and/or attached an Office communication concerning this application or proceeding.

## Office Action Summary

Application No.

10/022,466

Applicant(s)

GILMAN ET AL.

Examiner

Anthony D Stashick

Art Unit

3728

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133).
- Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 25 November 2003.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-12 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-12 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 13 December 2001 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
- Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- 11) ☐ The proposed drawing correction filed on \_\_\_\_\_ is: a) ☐ approved b) ☐ disapproved by the Examiner.
- If approved, corrected drawings are required in reply to this Office action.
- 12) ☐ The oath or declaration is objected to by the Examiner.

### Priority under 35 U.S.C. §§ 119 and 120

- 13) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- \* See the attached detailed Office action for a list of the certified copies not received.
- 14) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. § 119(e) (to a provisional application).
- a) ☐ The translation of the foreign language provisional application has been received.
- 15) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. §§ 120 and/or 121.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)                             | 4) <input type="checkbox"/> Interview Summary (PTO-413) Paper No(s). _____  |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)         | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449) Paper No(s) _____ | 6) <input type="checkbox"/> Other:  |

**DETAILED ACTION**

**IMPROPER INCORPORATION BY REFERENCE**

1. The attempt to incorporate subject matter into this application by reference to Patent Number 5,492,943 is improper because an application cannot incorporate by reference a patent that has incorporated another patent or application by reference, which occurs in Patent 5,492,943

***Claim Rejections - 35 USC § 102***

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

2. Claims 1-3 and 5-12 are rejected under 35 U.S.C. 102(b) as being anticipated by Jensen 5,827,213. Jensen '213 discloses all the limitations as claimed including the following: an orthotic pad having a hydrocolloid-containing adhesive body 13; the body having oppositely-facing non-parallel first 13b and second 13a major surfaces; the body comprising soft, deformable and shape-recoverable adhesive material 13; the first surface being contoured to match the contour of the treatment area

Art Unit: 3728

(flexible and can bend to match area); the second surface being adapted to match a support surface upon which the treatment area would contact in the absence of the orthotic pad (see Figures 2 and 6 bottom of the wound dressing flat to match the contact surface with the ground); a protective layer 14 covering the second surface; the protective layer is uniform in thickness and conforms with the shape of the second surface (see Figure 2); the protective layer is flexible (see col. 3, lines 65-67); second surface is generally planar (see Figures 2 and 6); second surface is non-planar (generally planar can have some areas of non-planar as shown in Figure 6 sides of foot); first major surface being non-planar (see Figure 2, 13a); second surface non-parallel to the first surface (see Figure 6, sides of foot); protective covering (15) on second surface; differing thickness throughout an outline area (see Figure 2) .

***Claim Rejections - 35 USC § 103***

3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the

Art Unit: 3728

art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

4. Claim 4 is rejected under 35 U.S.C. 103(a) as being unpatentable over Jensen 5,827,213 as applied to claim 1 or 2 above in view of Official Notice. Jensen '213 as applied to claims 1 or 2 above discloses all the limitations of the claim except for the protective layer being a rigid material. Official Notice is taken that since the protective layer is only used to protect the adhesive layer until time of use and then it is discarded, the protective layer can be made of any material that would allow for its release from the adhesive layer but would protect the adhesive layer from tears, indentions and dirt. Therefore, the release layer Jensen '213 can be made of any material desired to perform the functions of protecting the adhesive layer from being torn, punctured or dirtied before use, including flexible, rigid or semi-rigid material.

#### ***Response to Arguments***

5. Applicant's arguments filed November 25, 2003 have been fully considered but they are not persuasive. Applicant argues that Jensen does not disclose "an orthotic pad with a second surface adapted to match a support surface upon which the treatment area would contact in the absence of the orthotic

pad...". This argument is not clearly understood. The pad of Jensen, as can clearly be seen in Figures 2 and 6, clearly shows the second side of the pad adapted to match the flat contact surface of the ground which the foot would contact if the orthotic is not present, thereby meeting the limitations of the claim.

### *Conclusion*

6. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event,

however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Telephone inquiries regarding the status of applications or other general questions, by persons entitled to the information, "should be directed to the group clerical personnel and not to the examiners. In as much as the official records and applications are located in the clerical section of the examining groups, the clerical personnel can readily provide status information without contacting the examiners", M.P.E.P. 203.08. The Group clerical receptionist number is (703) 308-1148.

If in receiving this Office Action it is apparent to applicant that certain documents are missing, e.g., copies of references cited, form PTO-1449, form PTO-892, etc., requests for copies of such papers or other general questions should be directed to Tech Center 3700 Customer Service at (703) 306-5648, email [CustomerService3700@uspto.gov](mailto:CustomerService3700@uspto.gov).

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Anthony D Stashick whose telephone number is 703-308-3876. The examiner can normally be reached on Monday through Thursday 8:00 am-4:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Mickey Yu can be reached on 703-308-2672. The fax phone number for the organization where this application or proceeding is assigned is (703) 872-9306.

Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to the receptionist whose telephone number is 703-305-1148.

Art Unit: 3728

Other helpful telephone numbers are listed for applicant's benefit.

Allowed Files & Publication	(703) 305-8322
Assignment Branch	(703) 308-9287
Certificates of Correction	(703) 305-8309
Drawing Corrections/Draftsman	(703) 305-8404/8335
Fee Increase Questions	(703) 305-5125
Intellectual Property Questions	(703) 305-8217
Petitions/Special Programs	(703) 305-9282
Terminal Disclaimers	(703) 305-8408
Informal Fax for 3728	(703) 308-7769

If the information desired is not provided above, or has been changed, please do not call the examiner (this is the latest information provided to him) but the general information help line below.

Information Help line	1-800-786-9199
Internet PTO-Home Page	<a href="http://www.uspto.gov/">http://www.uspto.gov/</a>



Anthony D Stashick  
Primary Examiner  
Art Unit 3728

ADS

February 9, 2004